

## CUTTING-EDGE LASER TECHNOLOGIES FOR EXPEDITED MEDICAL DIAGNOSTICS AND TREATMENT

**Taranenko Tatyana Viktorovna**

Clinical Resident of Pediatric dentistry

**Tursunov Begzod Sherzodovich**

PhD, Department of Pediatric dentistry

Samarkand State Medical University

**Abstract:** The use of advanced laser technology for fast medical diagnosis and treatment is gaining prominence in medicine and dentistry, facilitated by innovative software solutions for Raman and fluorescence analysis.

The use of broadband, high-energy laser radiation fulfils two functions: firstly, as a therapeutic instrument [1, 2, 3, 4]; and secondly, as a means to augment the Raman luminescent diagnostic potential of a biological specimen [5] (specifically, in the context of biophotometry).

The methodology relies on the analysis of recorded spectral characteristics from Raman and/or luminescence fluorescence diagnostics (RLFD) of the subjects under investigation (microbial metabolites, cells, tissues, and biological fluids in both healthy and pathological conditions) to diagnose diseases and processes of microbial, metabolic, neoplastic, or other origins [6].

Our objective is to provide empirical data demonstrating that dental practice may gain advantages from the use of Raman-fluorescence medical technologies and laser hardware-software systems.

Constraints of Conventional Dental Diagnostics Oral health is essential for sustaining internal equilibrium amid internal and external pressures. Analysing dental health indicators enables the precise identification of pathological alterations, including oral infections, enamel mineralisation issues, and other prevalent dental disorders. Notwithstanding the importance of these attributes, dentistry now lacks the capability to effectively quantify them [1, 2, 3, 4].

Challenges associated with conventional diagnosis. Numerous impediments hinder conventional diagnostic methods. Although visual examination is essential for diagnosis, it must be undertaken with care owing to its inherent subjectivity and potential for varied interpretations. Multiple factors may affect the precision of a diagnosis. These encompass: Workplace illumination; Characteristics of colour perception as identified by medical professionals;

The health status of the healthcare professional. Disregarding these factors in conventional evaluation techniques results in substantial inaccuracies, sometimes reaching 30–40% [1, 2, 3, 4, 5, 6, 7, 8].

Employing Raman fluorescence for objective evaluation An important need exists to surmount the existing constraints in dental diagnosis methodologies. Raman fluorescence technology is a potential innovation that provides novel opportunities for evaluating oral health. Its utilisation permits: Perform a quantitative evaluation of the state of oral tissues.

Minimise the impact of subjectivity in diagnosis; Deliver operational outcomes [4, 5, 6, 7, 8].

This technology can assist in many areas, including: Identifying the initiation of demineralisation; Measurement of microbiological contamination levels. Evaluation of the current efficacy of the continuing therapy.

This methodology is optimal for ongoing condition monitoring, facilitates the standardisation of the diagnostic process, and is non-invasive, representing its three

primary benefits. This technique may enhance diagnostic accuracy by 25–30%, decrease examination duration, and facilitate the early identification of illnesses prior to symptom manifestation.

**Prospective pathways for expansion** The trajectory of future research in this domain will be dictated by the following variables: Development of hierarchical diagnostic methodologies; Developing computerised standards for evaluating oral health;

The use of technology that aid physicians in decision-making.

**Conclusion:** These novel methodologies provide a framework for the incorporation of evidence-based practice into standard clinical procedures and individualised dentistry. This ultimately enhances the quality of dental treatment.

### Reference:

1. Karaseva, V. V. Otsenka stomatologicheskogo statusa, kontrolya individual'noi gigieny i kachestva zhizni patsientov s defektami chelyustei v sochetanii s medikamentozno assotsirovannym osteonekrozom / V. V. Karaseva, T. M. Elovikova, A. S. Koshcheev // Stomatologiya. – 2020. – T. 99, № 5. – S. 80–86.
2. Otsenka stomatologicheskogo gigienicheskogo statusa studentov razlichnykh fakul'tetov Pervogo MGIMU im. I.M. Sechenova na osnovanii anketirovaniya i dannykh ob"ektivnogo obsledovaniya / I. M. Makeeva [i dr.] // Ros. stomatol. zhurn. – 2017. – T. 21, № 6. – S. 337–339. – DOI: 10.18821/1728-2802-2017-21-6-337-339.
3. Shirshova, N. E. Metodicheskie aspekty otsenki sostoyaniya gigieny polosti rta u lits molodogo vozrasta / N. E. Shirshova, V. R. Teslenko, O. S. Gileva // Perm. med. zhurn. – 2006. – № 6. – S. 107–112. – URL: <https://cyberleninka.ru/article/n/metodicheskie-aspekty-otsenki-sostoyaniya-gigieny-polosti-rta-u-lits-molodogo-vozrasta> (data obrashcheniya: 04.05.2025).

4. Otsenka urovnya gigieny polosti rta i sostoyaniya tkanei parodonta u internov i ordinatorov kafedry terapevticheskoi stomatologii Pervogo MGIMU im. I.M. Sechenova / O. I. Admakin [i dr.] // Sechenov. vestn. – 2015. – № 3 (21). – S. 27–30. – URL: <https://cyberleninka.ru/article/n/otsenka-urovnya-gigieny-polosti-rta-i-sostoyaniya-tkaney-parodonta-u-internov-i-ordinatorov-kafedry-terapevticheskoy-stomatologii> (data obrashcheniya: 04.05.2025).
5. Problemnye voprosy otsenki gigienicheskogo sostoyaniya polosti rta i ikh klinicheskoe reshenie / M. T. Aleksandrov [i dr.] // Stomatologiya. – 2020. – T. 99, № 4. – S. 21–26.
6. Primenenie lazernoi fluorestsentsii dlya otsenki gigienicheskogo sostoyaniya polosti rta / M. T. Aleksandrov [i dr.] // Vestn. RAMN. – 2003. – S. 39–44.
7. Lazernye Raman-fluorestsentnye meditsinskie tekhnologii v stomatologii: ot eksperimenta k klinike / M. T. Aleksandrov. – Moskva : KnigIzdat, 2020. – 160 s.
8. Stomatologiya (KhKh vek) : uchebnik dlya vuzov / M. T. Aleksandrov, N. N. Bazhanov [i dr.]. – M. : GOETAR-Media, 2008. – 415 s.